Centrix Incorporated
770 River Road, Shelton, Connecticut 06484 USA
www.centrixdental.com

Centrix Gun & Syringe Cleaning & Sterilization

Centrix, Inc. has validated the following methods for cleaning and sterilizing Centrix brand guns and syringes:

Cleaning

- Ultrasonic: 122°F (50°C) for 15-20 minutes
- Manual: Hand-scrub using a soft bristle brush in a mild detergent solution of approximately 10:1 water to detergent.

For both methods, rinse the gun or syringe under warm, sterile water, shake off the excess water, and then stand vertically to dry.

Sterilization

Gravity, steam sterilization can be performed under the following parameters:

- 121°C, 15 minutes, wrapped
- 121°C, 10 minutes, unwrapped
- 132°C, 10 minutes, wrapped
- 132°C, 3 minutes, unwrapped

If the above parameters are not possible, follow the specific cleaning and sterilization guidelines, including temperature, times, and drying cycles, as listed by your specific equipment manufacturers.

To obtain additional information about reprocessing these Centrix devices contact:

Centrix Customer Service
csorders@centrixdental.com
800-235-5862 option 2
https://www.centrixdental.com/products/delivery-systems/c-r-syringes-guns

Sincerely,



Tim White
Regulatory Affairs Manager

2026.11.10